

EN 60079-26

COMPANY WITH QUALITY SYSTEM CERTIFIED BY DNV GL SYSTEM CEA VAT Nr. IT 03359340548 Sir Safety System S.p.A. Single-member company export@sirsafety.com Wholly-owned subsidiary of Sir Holding Srl Viale Europa 74 www.sirsafety.com = ISO 9001 = Via dei Fornaciai, 9 - Santa Maria degli Angeli 20090 Cusago Milano DNV.GL 06081 Assisi - Perugia - Italy Tel. +39 02.90394575 (R.A.) Tel.+39 075.8043737 (R.A.) - Fax +39 075.8048686 Fax +39 02.90394441 ISO 9001 Name: Gas Alert Extreme ETO Code: 81567 Gas Detectors - BW Description Gas Alert Extreme ETO features: - One sensor. - Full function self-test of sensor, battery status, circuit integrity and audio/visual alarms. It constantly shows the gas concentration on the display. - Low, high, TWA and STEL alarms with adjustable settings. - Acoustic (95dB), visual and vibrating alarm. Impact and water resistant casing. Also available in data logging version. - Compact 2.8 x 5.0 x 9.5 cm. - Lightweight 82 g. Gas detector: Ethylene Oxide. The product is designed and manufactured to comply with the essential health and safety requirements of Directive 89/686/EEC, and to conform to the requirements of EN 60079. It is classified as PPE Cat. III. Available Sizes Category Colours III Yellow Quantity/Carton Important Information Recommended for Suitable to detect the presence of Ethylene Oxide. Make sure the product is in good condition, undamaged and clean before use. Once you become aware of a damage, you should 1 immediately replace it. Dexterity EC Standards Storage and cleaning EN 60079-0 Keep in a cool dry place, far from heat sources. EN 60079-11