



Sir Safety System S.p.A. Single-member company Wholly-owned subsidiary of Sir Holding Srl Via dei Fornaciai, 9 - Santa Maria degli Angeli 06081 Assisi - Perugia - Italy Tel.+39 075.8043737 (R.A.) - Fax +39 075.8048686

VAT Nr. IT 03359340548 Viale Europa 74 20090 Cusago Milano Tel. +39 02.90394575 (R.A.) Fax +39 02.90394441

export@sirsafety.com www.sirsafety.com

COMPANY WITH QUALITY SYSTEM CERTIFIED BY DNV GL



Name: Gas Alert Max XT II 4 Gas (LEL, O2, CO, H2S)

Code: 81610

Gas Detectors - BW

Description


Gas Alert Max XT II 4 Gas detects up to 4 gases simultaneously (Fuels, Oxygen, Sulphuric Acid, Carbon Monoxide). It features:

- It constantly shows the gas concentration on the display.
- Full function self-test of sensor, battery status, circuit integrity and audio/visual alarms.
- Integrated sampling pump with a 3 m hose.

for gas detection in confined spaces.

- Impact and water resistant casing.
- Avanced microprocessors, and easily plug in/replaceable catalytic and electrochemical sensors.
- Low, high, TWA (Time Weighted Average),
- STEL (Short Term Exposure Limit) alarm for toxic gases and OL (Over Limit) alarm for combustible gases, low battery.
- Acoustic (95dB), visual and vibrating alarm.
- Rechargeable batteries.
- Compact 13.1 x 7.0 x 5.2 cm.
- Weight 328 g.

The product is designed and manufactured to comply with the essential to the requirements of EN 60079.



Jacogory	00.00.0	7 transacto elecc
III	Yellow	
Quantity/Carton	Important Information	Recommended for
1	Make sure the product is in good condition, undamaged and clean before use. Once you become aware of a damage, you should immediately replace it.	Suitable to detect the presence of Fuels, Oxygen, Sulphuric Acid, Carbon Monoxide.
Dexterity	EC Standards	Storage and cleaning
	EN 60079-0	Keep in a cool dry place, far from heat sources.
	EN 60079-11	
	EN 60079-26	